


PILE DYNAMICS, INC.

GAGE 492H

28-Jul-93



PDA Cal Factor = 107.2 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 372.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
76.	.72
163.	1.52
232.	2.17
306.	2.87
377.	3.53
448.	4.18
524.	4.89
603.	5.62
673.	6.28
747.	6.98