


PILE DYNAMICS, INC.

GAGE 516H

10-AUG-93



PDA Cal Factor = 102.7 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 356.6 ME/mV/V

Strain ME	Gage Volts
0.	.00
72.	.70
144.	1.39
219.	2.11
288.	2.79
361.	3.49
432.	4.18
506.	4.91
581.	5.64
647.	6.29
716.	6.98