


PILE DYNAMICS, INC.

GAGE 521H

06-AUG-93



PDA Cal Factor = 103.2 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 358.3 ME/mV/V

Strain ME	Gage Volts
0.	.00
74.	.72
142.	1.37
221.	2.13
292.	2.81
366.	3.52
434.	4.19
505.	4.87
577.	5.57
648.	6.27
719.	6.98