


PILE DYNAMICS, INC.

GAGE 522H

06-Aug-93



PDA Cal Factor = 107.7 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 374.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
76.	.70
153.	1.41
231.	2.12
305.	2.81
379.	3.49
463.	4.27
533.	4.92
602.	5.56
675.	6.26
752.	7.00