


PILE DYNAMICS, INC.

GAGE 527H

06-Aug-93



PDA Cal Factor = 111.7 ME/V
 Ref. Shunt (50.4KQ) = 5.0 V
 General Gage Factor = 387.8 ME/mV/V

Strain ME	Gage Volts
0.	.00
81.	.72
162.	1.44
239.	2.12
315.	2.79
396.	3.52
477.	4.24
556.	4.95
632.	5.64
704.	6.29
782.	7.02