


PILE DYNAMICS, INC.

GAGE

564H

05-Jan-94



PDA Cal Factor = 115.6 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 401.3 ME/mV/V

Strain ME	Gage Volts
0.	.00
90.	.77
167.	1.46
238.	2.10
326.	2.86
402.	3.53
485.	4.23
571.	4.99
644.	5.60
731.	6.33
814.	7.03