


PILE DYNAMICS, INC.

GAGE 565H

11-AUG-93



PDA Cal Factor = 107.3 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 372.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
74.	.69
149.	1.39
226.	2.10
302.	2.80
380.	3.52
453.	4.20
525.	4.88
603.	5.61
678.	6.32
749.	6.99