


PILE DYNAMICS, INC.

GAGE 568H

10-Aug-93



PDA Cal Factor = 105.3 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 365.7 ME/mV/V

Strain ME	Gage Volts
0.	.00
65.	.62
139.	1.32
205.	1.94
288.	2.72
367.	3.47
430.	4.06
510.	4.83
579.	5.48
647.	6.14
728.	6.93