


PILE DYNAMICS, INC.

GAGE 571H

10-Aug-93



PDA Cal Factor = 101.2 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 351.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
82.	.81
142.	1.39
211.	2.07
286.	2.81
348.	3.43
436.	4.28
493.	4.85
569.	5.61
628.	6.20
707.	7.00