


PILE DYNAMICS, INC.

GAGE

572H

10-Aug-93



PDA Cal Factor = 108.3 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 376.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
82.	.76
155.	1.41
226.	2.07
299.	2.74
396.	3.41
456.	4.18
530.	4.86
610.	5.62
679.	6.25
765.	7.08