


PILE DYNAMICS, INC.

GAGE 573H

10-Aug-93



PDA Cal Factor = 103.5 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 359.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
95.	.92
136.	1.32
224.	2.15
300.	2.88
364.	3.49
444.	4.27
515.	4.96
579.	5.59
640.	6.18
732.	7.09