


PILE DYNAMICS, INC.

GAGE 574H

10-Aug-93

PDA Cal Factor = 101.8 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 353.6 ME/mV/V



Strain ME	Gage Volts
0.	.00
89.	.87
142.	1.40
221.	2.15
293.	2.87
359.	3.52
426.	4.17
500.	4.89
589.	5.77
643.	6.31
717.	7.06