


PILE DYNAMICS, INC.

GAGE 577H

10-Aug-93



PDA Cal Factor = 105.7 ME/V
 Ref. Shunt (60.4K) = 5.0 V

General Gage Factor = 366.9 ME/mV/V

Strain ME	Gage Volts
0.	.00
65.	.62
140.	1.31
211.	1.99
287.	2.70
373.	3.51
440.	4.14
501.	4.73
579.	5.47
669.	6.32
733.	6.96