


PILE DYNAMICS, INC.

GAGE

582H

19-Aug-93



PDA Cal Factor = 109.6 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 380.6 ME/mV/V

Strain ME	Gage Volts
0.	.00
89.	.81
144.	1.31
232.	2.10
313.	2.83
397.	3.59
459.	4.16
533.	4.84
614.	5.59
694.	6.32
778.	7.12