


PILE DYNAMICS, INC.

GAGE 583H

19-Aug-93



PDA Cal Factor = 100.2 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 347.8 ME/mV/V

Strain ME	Gage Volts
0.	.00
68.	.67
137.	1.36
227.	2.25
272.	2.70
375.	3.72
447.	4.44
502.	4.99
574.	5.72
623.	6.22
706.	7.06