


PILE DYNAMICS, INC.

GAGE 584H

19-AUG-93



PDA Cal Factor = 112.1 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 389.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
59.	.54
139.	1.24
257.	2.27
310.	2.74
392.	3.47
474.	4.21
543.	4.81
614.	5.46
685.	6.11
775.	6.93