


PILE DYNAMICS, INC.

GAGE 602H

21-Feb-94



PDA Cal Factor = 106.8 ME/V
 Shunt (60.4K) = 5.0 V
 General Factor = 370.7 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0.	.00
70.	.67
145.	1.38
221.	2.10
293.	2.78
362.	3.43
442.	4.18
512.	4.82
585.	5.51
670.	6.28
739.	6.93

Calibrated by _____