


PILE DYNAMICS, INC.

GAGE 605H

01-Sep-93



PDA Cal Factor = 101.1 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 351.2 ME/mV/V

Strain ME	Gage Volts
0.	.00
60.	.60
133.	1.30
207.	2.03
268.	2.62
343.	3.36
416.	4.09
487.	4.79
554.	5.46
624.	6.16
695.	6.89