


PILE DYNAMICS, INC.

GAGE 606H

01-Sep-93



PDA Cal Factor = 104.4 ME/V
 Ref. Shunt (60.4KQ) = 5.0 V
 General Gage Factor = 362.3 ME/mV/V

Strain ME	Gage Volts
0.	.00
69.	.65
139.	1.31
206.	1.95
292.	2.77
365.	3.46
427.	4.06
506.	4.81
574.	5.47
650.	6.23
713.	6.84