


PILE DYNAMICS, INC.

GAGE 607H

01-Sep-93



PDA Cal Factor = 102.9 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 357.1 ME/mV/V

Strain ME	Gage Volts
0.	.00
81.	.78
155.	1.49
225.	2.16
288.	2.77
366.	3.53
434.	4.18
505.	4.87
575.	5.56
651.	6.32
720.	7.01