


PILE DYNAMICS, INC.

GAGE 609H

01-Sep-93



PDA Cal Factor = 101.8 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 353.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
66.	.64
131.	1.28
209.	2.03
275.	2.67
351.	3.42
417.	4.07
489.	4.77
563.	5.50
631.	6.19
704.	6.93