


PILE DYNAMICS, INC.

GAGE 610H

01-Sep-93



PDA Cal Factor = 101.7 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 353.0 ME/mV/V

Strain ME	Gage Volts
0.	.00
81.	.79
156.	1.52
226.	2.20
294.	2.86
362.	3.53
426.	4.16
497.	4.86
575.	5.63
644.	6.33
709.	6.98