


PILE DYNAMICS, INC.

GAGE 611H

01-Sep-93



PDA Cal Factor = 105.4 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 365.9 ME/mV/V

Strain ME	Gage Volts
0.	.00
68.	.63
157.	1.47
226.	2.12
299.	2.80
371.	3.48
441.	4.14
514.	4.84
587.	5.54
659.	6.24
736.	6.99