


PILE DYNAMICS, INC.

GAGE 612H

01-Sep-93



PDA Cal Factor = 102.4 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 355.5 ME/mV/V

Strain ME	Gage Volts
0.	.00
66.	.65
144.	1.38
203.	1.96
285.	2.76
359.	3.48
418.	4.04
502.	4.87
580.	5.65
631.	6.15
710.	6.94