


PILE DYNAMICS, INC.

GAGE 613H

01-Sep-93



PDA Cal Factor = 100.9 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 350.4 ME/mV/V

Strain ME	Gage Volts
0.	.00
49.	.48
136.	1.33
210.	2.06
268.	2.63
340.	3.33
422.	4.15
482.	4.74
556.	5.48
617.	6.11
685.	6.80