


PILE DYNAMICS, INC.

GAGE

626H

22-Sep-93



PDA Cal Factor = 161.0 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 558.9 ME/mV/V

Strain ME	Gage Volts
0.	.00
75.	.47
166.	1.03
246.	1.52
323.	2.00
403.	2.49
483.	2.98
570.	3.52
650.	4.03
733.	4.56
812.	5.05