


PILE DYNAMICS, INC.

GAGE

627H

22-Sep-93



PDA Cal Factor = 164.3 ME/V
 Ref. Shunt (60.4K) = 5.0 V
 General Gage Factor = 570.6 ME/mV/V

Strain ME	Gage Volts
0.	.00
89.	.55
169.	1.03
251.	1.52
340.	2.06
423.	2.56
500.	3.03
582.	3.53
662.	4.02
737.	4.48
816.	4.98