


PILE DYNAMICS, INC.

GAGE 628H

22-Sep-93



PDA Cal Factor = 171.9 ME/V
 Ref. Shunt (60.4KΩ) = 5.0 V
 General Gage Factor = 596.7 ME/mV/V

Strain ME	Gage Volts
0.	.00
92.	.53
172.	.98
275.	1.58
366.	2.10
442.	2.54
527.	3.02
606.	3.49
698.	4.04
780.	4.54
858.	5.00