


Pile Dynamics, Inc. Transducer 801H



PDA Cal Factor = 109.8 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 381.3 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
74	.69
176	1.64
234	2.17
321	2.97
377	3.48
452	4.16
526	4.83
603	5.52
699	6.38
766	6.98

Calibrated by: J. E. Schubert
 Calibrated on: 15-May-96