


Pile Dynamics, Inc. Transducer 807H



PDA Cal Factor = 110.6 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 384.0 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
68	.62
154	1.42
227	2.08
301	2.76
382	3.48
467	4.26
531	4.83
616	5.59
675	6.12
762	6.89

Calibrated by: *J. B. Smith*
 Calibrated on: 11-Sep-96