


Pile Dynamics, Inc. Transducer 809H



PDA Cal Factor = 107.0 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 371.4 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
65	.61
164	1.55
218	2.06
317	2.99
364	3.42
474	4.44
512	4.80
611	5.71
676	6.33
750	7.02

Calibrated by: *J. Embert*
 Calibrated on: 10-Feb-96