


Pile Dynamics Inc Transducer 5114



PDA Cal Factor = 97.5 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 338.7 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
186	1.92
195	2.02
223	2.29
307	3.15
395	4.06
422	4.33
498	5.10
590	6.05
653	6.70
703	7.22

Calibrated by: J. D. Dettlaff

Calibrated on: 29-Dec-00