


Pile Dynamics Inc Transducer 5115



PDA Cal Factor = 96.3 ME/V
 Shunt (60.4KΩ) = 5.0 V
 General Factor = 334.2 ME/mV/V
 Traceable to N.I.S.T.

Strain ME	Gage V
0	.00
199	2.08
205	2.14
204	2.13
322	3.35
410	4.27
410	4.28
508	5.28
629	6.54
633	6.59
669	6.96

Calibrated by: J. J. Cantelmo
 Calibrated on: 29-Dec-00