


Cycle 1
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	1.12926095	-0.064884649
3	2.02438413	-0.116182994
4	3.23250302	-0.184650357
5	4.14868095	-0.236754416
6	5.08581626	-0.289578838
7	6.06993513	-0.345417254
8	7.10981047	-0.404065905
9	8.23312559	-0.466908569
10	8.97433436	-0.509772048
11	10.0222034	-0.573330448

Serial Number 30303
Force Calibration (lb/V) -17.5866111
Offset -0.00984148
Correlation -0.99997346




Cycle 2
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	0.85564545	-0.047384047
3	1.83235615	-0.102103924
4	3.06601379	-0.170410251
5	4.00373393	-0.222296392

6	4.98736535	-0.275860157
7	5.94692024	-0.329836268
8	7.00463372	-0.389622351
9	8.12911856	-0.452062313
10	8.93729351	-0.494196072
11	9.95231309	-0.560171444

Serial Number 30303
 Force Calibration (kips/mV/V) -17.8960448
 Offset 0.02408496
 Correlation -0.99990418



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	1.15285015	-0.066121633
3	2.07536411	-0.118969711
4	2.96756293	-0.169751683
5	3.8917339	-0.222983718
6	4.84982657	-0.276945604
7	6.12929808	-0.350133322
8	6.85530028	-0.391371368
9	8.17444496	-0.466530682
10	8.93690355	-0.509427476
11	10.0109937	-0.570844095

