


Cycle 1
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	0.86179133	-0.047142522
3	1.99690695	-0.11215546
4	2.90392697	-0.163548847
5	3.90462182	-0.219809128
6	4.99538485	-0.280633029
7	6.07513286	-0.342779146
8	6.82092606	-0.386178159
9	7.99464129	-0.45154212
10	8.75944233	-0.496533498
11	9.83519358	-0.559006845

Serial Number 30383
Force Calibration (lb/V) -17.571657
Offset 0.03904926
Correlation -0.99998593




Cycle 2
Sample

	Force (kips)	Bridge (mV/V)
1	0	0
2	1.0667848	-0.060430906
3	2.17509432	-0.122165097
4	3.08113956	-0.17473424
5	4.043916	-0.230483284


6	5.10972503	-0.29121779
7	6.24903122	-0.355535046
8	6.95651612	-0.397958333
9	8.10098839	-0.46513811
10	8.87271027	-0.511551749
11	10.0283923	-0.582641408

Serial Number 30383
 Force Calibration (kips/mV/V) -17.2072659
 Offset 0.07611836
 Correlation -0.99991941



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	0.81851169	-0.04674028
3	1.92126507	-0.107887048
4	2.84524603	-0.159603361
5	3.82527585	-0.21404465
6	4.87656088	-0.273637285
7	5.99286261	-0.337570929
8	6.75006057	-0.37952996
9	7.89765209	-0.444534805
10	9.11562163	-0.513458937
11	9.84689074	-0.556114818

Serial Number 30383
Force Calibration (kips/mV/V) -17.7250127
Offset 0.01278199
Correlation -0.99998963



Serial Number 30383
Force Calibration (kips/mV/V) -17.5013119

Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3500.26237

Calibrated by: Sean Bonno
Calibrated Date: 7/11/2024

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

