


Cycle 1
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 1.10567548	-0.061889772
3 2.27821971	-0.129845368
4 3.27559875	-0.188128956
5 3.97557716	-0.228850256
6 5.04547882	-0.289934631
7 6.17630347	-0.35606601
8 6.96011202	-0.402128193
9 8.1492282	-0.470131603
10 9.31387827	-0.540025955
11 10.0446604	-0.583169158

Serial Number 30398
Force Calibration (lb/V) -17.1724149
Offset 0.05114662
Correlation -0.99998975




Cycle 2
Sample

Force (kips)	Bridge (mV/V)
1 0	0
2 0.86822282	-0.049952537
3 2.10490656	-0.12034866
4 3.1059897	-0.179456817
5 3.83228676	-0.219595236


6	4.9424462	-0.283229142
7	6.04792697	-0.348199497
8	6.82832385	-0.393896788
9	8.05321394	-0.463032806
10	8.81450591	-0.509028769
11	10.0148328	-0.5750752

Serial Number 30398
 Force Calibration (kips/mV/V) -17.3626269
 Offset 0.00624862
 Correlation -0.99998343



Sample	Cycle 3	
	Force (kips)	Bridge (mV/V)
1	0	0
2	1.1155206	-0.06360999
3	1.99290613	-0.113420398
4	2.93355387	-0.167960253
5	3.98464247	-0.22715375
6	5.10513442	-0.290370633
7	6.2181209	-0.355089827
8	7.01430899	-0.400706556
9	8.22564988	-0.470136387
10	8.99512992	-0.515454685
11	9.86803289	-0.563242016

Serial Number 30398
Force Calibration (kips/mV/V) -17.481233
Offset 0.00874091
Correlation -0.99999165



Serial Number 30398
Force Calibration (kips/mV/V) -17.3387583

Squid Gain 50
Squid Excitation 5
Force Calibration (kips/V) -3467.75165

Calibrated by: Sean Bonner
Calibrated Date: 1/22/2025

Pile Dynamics Inc
30725 Aurora Rd
Solon, OH 44139

Traceable to N.I.S.T.

