




Pile Dynamics, Inc.

Transducer E756

Strain ($\mu\epsilon$)



PDA Cal Factor (5.0 V)

94.3 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
155	0.75
288	1.45
418	2.19
525	2.76
678	3.56
819	4.32
933	4.93
1061	5.62
1187	6.27
1353	7.10

Shunt (60.4 K Ω)

2.5 V

General Factor


327.3 $\mu\epsilon/mV/V$

Traceable to N.I.S.T.

Strain Transducer Calibrator System 2009 Version 1.3

Calibrated by:

Calibrated on:


18-May-2010