




# Pile Dynamics, Inc.

## Transducer E776

Strain ( $\mu\epsilon$ )



PDA Cal Factor (5.0 V)

93.8 $\mu\epsilon/V$

Applied Strain ( $\mu\epsilon$ )	Transducer Output (Volts)
0	0.00
131	0.71
259	1.40
399	2.14
523	2.81
656	3.53
783	4.20
920	4.93
1058	5.65
1184	6.34
1315	7.00

Shunt (60.4 K $\Omega$ )  
General Factor

2.5 V  
325.6 $\mu\epsilon/mV/V$

Traceable to N.I.S.T.

Strain Transducer Calibrator System 2009 Version 1.3

Calibrated by: Kosta Kosta

Calibrated on: 21-Dec-2009