




Pile Dynamics, Inc.

Pile Dynamics, Inc.

Transducer E790

Strain ($\mu\epsilon$)



Traceable to N.I.S.T.

Strain Transducer Calibrator System 2011 Version 1.5

PDA Cal Factor (5.0 V) 97.6 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
165	0.80
279	1.41
413	2.12
551	2.83
683	3.52
832	4.27
958	4.92
1095	5.62
1245	6.36
1380	7.03

Shunt (60.4 K Ω)

2.5 V

General Factor

338.9 $\mu\epsilon/mVV$

Calibrated by:

Calibrated on:

09-Mar-2012

Smart Sensor? Yes No Smart Chip

Programmed by: _____ Date: _____ CRC value _____