




Pile Dynamics, Inc.

Pile Dynamics, Inc.

Transducer G831

Strain ($\mu\epsilon$)



PDA Cal Factor (5.0 V)

96.7 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
139	0.70
273	1.40
409	2.11
541	2.80
675	3.50
811	4.20
945	4.90
1081	5.59
1220	6.31
1359	7.00

Shunt (60.4 K Ω)

2.5 V

General Factor

335.7 $\mu\epsilon/mV/V$

Traceable to N.I.S.T.

Strain Transducer Calibrator System 2009 Version 1.3

Calibrated by: Kathleen Hillman

Calibrated on: 07-Jul-2011

Smart Sensor? Yes No Smart Chip Programmed by: VW on: 7/7/11 CRC value 05E0