


Pile Dynamics, Inc.

Transducer G856

Strain ($\mu\epsilon$)



PDA Cal Factor (5.0 V)

92.3 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
129	0.69
258	1.40
387	2.11
518	2.82
640	3.49
772	4.20
899	4.89
1037	5.63
1166	6.31
1300	7.02

Shunt (60.4 K Ω)

2.5 V

General Factor

320.6 $\mu\epsilon/mV/V$

Traceable to N.I.S.T.

Strain Transducer Calibrator System 2009 Version 1.3

Calibrated by:

Kurtis Holden

Calibrated on:

05-Jul-2011

Smart Sensor? Yes No Smart Chip Programmed by: VW on: 7/5/11 CRC value 82F8