




Pile Dynamics, Inc.

Pile Dynamics, Inc.

Transducer G863

Strain ($\mu\epsilon$)



PDA Cal Factor (5.0 V)

94.9 $\mu\epsilon/V$

Applied Strain ($\mu\epsilon$)	Transducer Output (Volts)
0	0.00
133	0.69
267	1.40
396	2.11
525	2.79
656	3.50
792	4.19
925	4.90
1066	5.62
1199	6.32
1330	6.99

Shunt (60.4 K Ω)

2.5 V

General Factor

329.5 $\mu\epsilon/mV/V$

Traceable to N.I.S.T.

Strain Transducer Calibrator System 2009 Version 1.3

Calibrated by: Karlus Hutter

Calibrated on: 05-Jul-2011

Smart Sensor? Yes No Smart Chip Programmed by: VW on: 7/5/11 CRC value EC05