
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 040D
PDI Gage Factor: 94.3
Mean Linear Correlation Coefficient: 0.999936

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1765.1	-5.394	551.2	1.710
-1279.5	-3.903	809.3	2.531
-794.0	-2.423	1069.6	3.339
-439.6	-1.327	1349.5	4.179
-131.2	-0.384	1627.3	5.010
245.0	0.778	2014.4	6.119

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kelly Beal Signature: Kelly Beal

PDI Gage: 040D

Calibration Date: OCT 2 2 2026

Calibration valid 2 years

Pile Dynamics, Inc.