
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 191D
PDI Gage Factor: 93.1
Mean Linear Correlation Coefficient: 0.999975

Table 1: Representative Calibration Data

Applied strain (με)	Gage Output (mV/Vexc)	Applied strain (με)	Gage Output (mV/Vexc)
-2064.7	-6.345	669.3	2.143
-1489.5	-4.573	984.3	3.138
-909.9	-2.779	1318.9	4.181
-505.2	-1.515	1664.5	5.216
-78.7	-0.215	2001.3	6.234
341.2	1.113		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kelly Beal

Signature: Kelly Beal

PDI Gage: 191D

Calibration Date: OCT 15 2026

Calibration valid 2 years

Pile Dynamics, Inc.