Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 656B
PDI Gage Factor: 92.8
Mean Linear Correlation Coefficient: 0.999953

Table 1: Representative Calibration Data

Applied strain (μE)	Gage Output (mV/Vexc)	Applied strain (μE)	Gage Output (mV/Vexc)
-1975.1	-6.173	662.7	2.098
-1456.7	-4.550	1017.1	3.180
-938.3	-2.947	1338.6	4.163
-524.9	-1.635	1653.5	5.139
-164.0	-0.536	2014.4	6.201
242.8	0.772		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kelly Beal Signature: Kelly Beal

PDI Gage: 656B

Calibration Date: OCT 19 2026

Calibration valid 2 years

Pile Dynamics, Inc.