
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 997C
PDI Gage Factor: 93.9
Mean Linear Correlation Coefficient: 0.999980

Table 1: Representative Calibration Data

Applied strain (με)	Gage Output (mV/Vexc)	Applied strain (με)	Gage Output (mV/Vexc)
-1791.3	-5.491	603.7	1.872
-1299.2	-4.003	879.3	2.744
-820.2	-2.503	1148.3	3.575
-406.8	-1.230	1430.4	4.420
-124.7	-0.369	1706.0	5.250
255.9	0.807	1986.0	6.094

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kelly Beal Signature: Kelly Beal

PDI Gage: 997C

Calibration Date: OCT 2 2 2026

Calibration valid 2 years

Pile Dynamics, Inc.