Certificate of Calibration
Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.

Serial No:	K15571	**Temperature:**	24.0 °C
Model:	KICHER PR	**Humidity:**	35%
Calibrated on:	Channel 3 on 8G 5161 LE		

Ref Acc 1:	85363! 1000 g's/volt	**Cal on:**	14Nov2025
Ref Acc 2:	85382! 958 g's/volt	**Cal on:**	24Nov2025

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

PDA CALIBRATION FACTOR


384.3 mv/5000g
(76.9 µv/g)
R^2: 0.999953 [Chip programmed]

Operator: William Johnson

Signed

Calibration valid 2 years

From: OCT 15 2026



Reference Velocity	S/N K15571 Velocity
m/s	m/s
1.642	1.632
2.089	2.079
2.452	2.450
2.864	2.861
3.190	3.188
3.494	3.485
3.799	3.800
4.079	4.084
4.348	4.356
4.686	4.693

Maximum Acceleration: 1015 g's