Accelerometer Calibration Certificate
Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.
Calibration performed on

Serial No:	10846	**Temperature:**	20.0 °C
Model:	PE	**Humidity:**	31%
Calibrated on:	Channel 4 on 8G 5161 LE		

Ref Acc 1:	65538! 1040 g's/volt	**Cal on:**	27Jan2020
Ref Acc 2:	63479! 1080 g's/volt	**Cal on:**	09Sep2020

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

PDA CALIBRATION FACTOR


1007.8 g's/volt

R^2: 0.999992 [Chip programmed]

Operator: William Johnson

Signed

PILE DYNAMICS, INC.
ORIGINAL COPY



Reference Velocity	S/N 10846 Velocity
m/s	m/s
0.530	0.526
0.891	0.889
1.094	1.092
1.834	1.841
2.193	2.198
2.733	2.731
3.018	3.016
3.435	3.431
3.639	3.639
4.120	4.121

Maximum Acceleration: 900 g's

NOV 18 2020

Version: 2020.30.170 0.08