
Certificate of Calibration

Pile Dynamics, Inc.
Transducer Model: PDI Strain Transducer
Serial Number: 182D
PDI Gage Factor: 145.5
Mean Linear Correlation Coefficient: 0.999904

Table 1: Representative Calibration Data

Applied strain (με)	Gage Output (mV/Vexc)	Applied strain (με)	Gage Output (mV/Vexc)
-2126.0	-4.255	734.9	1.496
-1513.6	-3.022	1069.6	2.140
-892.4	-1.757	1430.4	2.829
-413.4	-0.808	1784.8	3.511
-6.6	0.007	2005.7	3.946
413.4	0.843		

Calibration Curve



PDI Strain Transducer Calibration System (PDI STCS)

PDI STCS Serial Number:	1000HA
Firmware version number:	0.9.0.0
Transducer Gage Length:	3 inches (76.2mm)
Excitation Voltage for Calibration:	5.0 VDC

PDI certifies the above STCS instrument meets or exceeds published specifications and has been verified using standards and instruments whose accuracies are traceable to the National Institute of Standards and Technology (NIST), an accepted value of a natural physical constant or a ratio calibration technique.

Calibrated By: Kelly Beal Signature: Kelly Beal

PDI Gage: 182D

Calibration Date: NOV 17 2026

Calibration valid 2 years

Pile Dynamics, Inc.