Certificate of Calibration
Pile Dynamics, Inc.

Calibrated by Pile Dynamics, Inc.

Serial No:	87482	**Temperature:**	22.9 °C
Model:	PE	**Humidity:**	34%
Calibrated on:	Channel 3 on 8G 5161 LE		

Ref Acc 1:	85382! 958 g's/volt	**Cal on:**	24Nov2025
Ref Acc 2:	85363! 1000 g's/volt	**Cal on:**	14Nov2025

Reference accelerometer calibrations are traceable to the United States National Institute of Standards and Technology (NIST).

PDA CALIBRATION FACTOR

946.7 g's/volt

R^2: 0.999991 [Chip programmed]

Operator: William Johnson

Signed

Calibration valid 2 years

From: NOV 0 6 2026



Reference Velocity	S/N 87482 Velocity
m/s	m/s
1.695	1.697
2.166	2.167
2.447	2.443
2.877	2.879
3.283	3.278
3.532	3.531
3.825	3.821
4.102	4.105
4.335	4.336
4.625	4.628

Maximum Acceleration: 1045 g's